


INVITATION TO BID STATE OF LOUISIANA DIVISION OF ADMINISTRATION OFFICE OF STATE PURCHASING		BIDS WILL BE PUBLICLY OPENED: <div style="font-size: 1.2em; font-weight: bold;">MAY 27, 2010 10:00 AM</div>
<div style="border: 1px solid black; padding: 5px;"> =====> VENDOR NO. : SOLICITATION : 2239603 FILE NO. : R12015N OPENING DATE : 05/27/10 </div> <div style="border: 1px solid black; padding: 5px; margin-top: 10px;"> =====> VENDOR NAME AND ADDRESS </div> <div style="text-align: center; font-weight: bold; font-size: 1.1em;"> FILL IN VENDOR NUMBER (FEIN), NAME AND ADDRESS ABOVE, BEFORE SUBMITTING BID. </div>		PURCHASING AGENCY NO. : 107001 SEE NO. 8 BELOW. RETURN BID TO <div style="text-align: right; font-weight: bold;">10:00 AM</div> <div style="display: flex; justify-content: space-between;"> 2239603 05/27/10 R12015N </div> OFFICE OF STATE PURCHASING OFFICE OF STATE PURCHASING POST OFFICE BOX 94095 BATON ROUGE, LA 70804-9095
FY11 HEMATOLOGY ANALYZER RENT & SUPPLIES VILLA FELICIANA MEDICAL CENTER		<div style="font-size: 0.9em;"> BUYER : JANA BRAUD BUYER PHONE : (225) 342-8036 DATE ISSUED : 05/04/10 REQ. AGENCY : 320VFC FOLD HERE--> VILLA FELICIANA HOSPITAL AGENCY REQ. NO. : ISIS REQ. NO. : 1331566 VENDOR PHONE : FISCAL YEAR : 11 CLASS/SUBCLASS : 97917 SCHEDULED BEGIN DATE : 00/00/00 SCHEDULED END DATE : 00/00/00 T-NUMBER : </div>
TO BE COMPLETED BY VENDOR		
1. _____ PLEASE REMOVE FROM THIS COMMODITY CODE. 2. _____ DELIVERY WILL BE MADE IN THIS NUMBER OF DAYS AFTER RECEIPT OF ORDER. 3. _____ % CASH DISCOUNT FOR PROMPT PAYMENT IF MADE WITHIN THIRTY (30) DAYS. CASH DISCOUNTS FOR LESS THAN 30 DAYS OR LESS THAN 1% WILL BE ACCEPTED, BUT WILL NOT BE CONSIDERED IN DETERMINING AWARDS. ON INDEFINITE QUANTITY TERM CONTRACTS, CASH DISCOUNTS WILL BE ACCEPTED AND TAKEN BUT WILL NOT BE CONSIDERED IN DETERMINING AWARDS. 4. _____ BID BOND ATTACHED, _____ CERTIFIED CHECK ATTACHED, _____ OTHER, IF REQUIRED. 5. _____ BID REFERENCE NUMBER. (THIS NUMBER WILL APPEAR ON RESULTING ORDER OR CONTRACT).		
INSTRUCTIONS TO BIDDERS		
1. READ THE ENTIRE BID, INCLUDING ALL TERMS AND CONDITIONS AND SPECIFICATIONS. 2. ALL BID PRICES MUST BE TYPED OR WRITTEN IN INK. ANY CORRECTIONS, ERASURES OR OTHER FORMS OF ALTERATION TO UNIT PRICES SHOULD BE INITIALED BY THE BIDDER. 3. THIS BID IS TO BE MANUALLY SIGNED IN INK. FOLD HERE--> 4. BID PRICES SHALL INCLUDE DELIVERY OF ALL ITEMS F.O.B. DESTINATION OR AS OTHERWISE PROVIDED. BIDS CONTAINING "PAYMENT IN ADVANCE" OR "C.O.D" REQUIREMENTS MAY BE REJECTED. PAYMENT IS TO BE MADE WITHIN 30 DAYS AFTER RECEIPT OF PROPERLY EXECUTED INVOICE OR DELIVERY, WHICHEVER IS LATER. 5. AMOUNT OF BID BOND REQUIRED: _____ N/A _____ . 6. AMOUNT OF PERFORMANCE BOND, IF REQUIRED. _____ OR _____ 0% _____ OF BID. 7. DESIRED DELIVERY: _____ 120DAYS ARO _____ 8. TO ASSURE CONSIDERATION OF YOUR BID, ALL BIDS AND ADDENDA SHOULD BE RETURNED IN AN ENVELOPE OR PACKAGE CLEARLY MARKED WITH THE BID OPENING DATE AND THE BID NUMBER, OR SUBMITTED IN THE SPECIAL ENVELOPE IF FURNISHED FOR THAT PURPOSE. 9. BIDS SUBMITTED ARE SUBJECT TO PROVISIONS OF THE LAWS OF THE STATE OF LOUISIANA INCLUDING BUT NOT LIMITED TO L.R.S. 39:1551-1736; PURCHASING RULES AND REGULATIONS; EXECUTIVE ORDERS; STANDARD TERMS AND CONDITIONS; SPECIAL CONDITIONS; AND SPECIFICATIONS LISTED IN THIS SOLICITATION. 10. IMPORTANT: BY SIGNING THE BID, THE BIDDER CERTIFIES COMPLIANCE WITH ALL INSTRUCTIONS TO BIDDERS, TERMS, CONDITIONS AND SPECIFICATIONS, AND FURTHER CERTIFIES THAT THIS BID IS MADE WITHOUT COLLUSION OR FRAUD. THIS BID IS TO BE MANUALLY SIGNED IN INK BY A PERSON AUTHORIZED TO BIND THE VENDOR (SEE NO.30). ALL BID INFORMATION SHALL BE MADE WITH INK OR TYPEWRITTEN.		
VENDOR PHONE NUMBER: FAX NUMBER:	TITLE	DATE
SIGNATURE OF AUTHORIZED BIDDER - SEE NO. 30, PAGE 3. (MUST BE SIGNED)	NAME OF BIDDER (TYPED OR PRINTED)	

STANDARD TERMS & CONDITIONS		INVITATION TO BID	
NUMBER : 2239603 OPEN DATE: 05/27/10 TIME: 10:00 AM T-NUMBER :		BIDDER:	PAGE 2
<p>11 ADDRESS ALL INQUIRIES AND CORRESPONDENCE TO THE BUYER AT THE PHONE AND ADDRESS SHOWN ABOVE.</p> <p>12. CONFERENCE: NA NA NA</p> <p>13. BID FORMS. ALL WRITTEN BIDS, UNLESS OTHERWISE PROVIDED FOR, MUST BE SUBMITTED ON, AND IN ACCORDANCE WITH, FORMS PROVIDED, PROPERLY SIGNED (SEE NO. 30). BIDS SUBMITTED IN THE FOLLOWING MANNER WILL NOT BE ACCEPTED:</p> <ul style="list-style-type: none">A. BID CONTAINS NO SIGNATURE INDICATING INTENT TO BE BOUND;B. BID FILLED OUT IN PENCIL; ANDC. BID NOT SUBMITTED ON THE STATE'S STANDARD FORMS. <p>BIDS MUST BE RECEIVED AT THE ADDRESS SPECIFIED IN THE SOLICITATION PRIOR TO BID OPENING TIME IN ORDER TO BE CONSIDERED. TELEGRAPHIC AND FAX ALTERATIONS TO BIDS RECEIVED BEFORE BID OPENING TIME WILL BE CONSIDERED PROVIDED FORMAL BID AND WRITTEN ALTERATION HAVE BEEN RECEIVED AND TIME-STAMPED BEFORE BID OPENING TIME. ENTIRE BID SHOULD BE RETURNED, EXCEPT ITEM PAGES NOT BID.</p> <p>14. STANDARDS OF QUALITY. ANY PRODUCT OR SERVICE BID SHALL CONFORM TO ALL APPLICABLE FEDERAL AND STATE LAWS AND REGULATIONS AND THE SPECIFICATIONS CONTAINED IN THE SOLICITATION. UNLESS OTHERWISE SPECIFIED IN THE SOLICITATION, ANY MANUFACTURER'S NAME, TRADE NAME, BRAND NAME, OR CATALOG NUMBER USED IN THE SPECIFICATION IS FOR THE PURPOSE OF DESCRIBING THE STANDARD OF QUALITY, PERFORMANCE, AND CHARACTERISTICS DESIRED AND IS NOT INTENDED TO LIMIT OR RESTRICT COMPETITION. BIDDER MUST SPECIFY THE BRAND AND MODEL NUMBER OF THE PRODUCT OFFERED IN HIS BID. BIDS NOT SPECIFYING BRAND AND MODEL NUMBER SHALL BE CONSIDERED AS OFFERING THE EXACT PRODUCTS SPECIFIED IN THE SOLICITATION.</p> <p>15. DESCRIPTIVE INFORMATION. BIDDERS PROPOSING AN EQUIVALENT BRAND OR MODEL SHOULD SUBMIT WITH THE BID INFORMATION (SUCH AS ILLUSTRATIONS, DESCRIPTIVE LITERATURE, TECHNICAL DATA) SUFFICIENT FOR STATE OF LOUISIANA TO EVALUATE QUALITY, SUITABILITY, AND COMPLIANCE WITH THE SPECIFICATIONS IN THE SOLICITATION. FAILURE TO SUBMIT DESCRIPTIVE INFORMATION MAY CAUSE BID TO BE REJECTED. ANY CHANGE MADE TO A MANUFACTURER'S PUBLISHED SPECIFICATIONS SUBMITTED FOR A PRODUCT SHALL BE VERIFIABLE BY THE MANUFACTURER. IF ITEM(S) BID DO NOT FULLY COMPLY WITH SPECIFICATIONS (INCLUDING BRAND AND/OR PRODUCT NUMBER), BIDDER MUST STATE IN WHAT RESPECT ITEM(S) DEVIATE. FAILURE TO NOTE EXCEPTIONS ON THE BID FORM WILL NOT RELIEVE THE SUCCESSFUL BIDDER(S) FROM SUPPLYING THE ACTUAL PRODUCTS REQUESTED.</p> <p>16. BID OPENING. BIDDERS MAY ATTEND THE BID OPENING, BUT NO INFORMATION OR OPINIONS CONCERNING THE ULTIMATE CONTRACT AWARD WILL BE GIVEN AT THE BID OPENING OR DURING THE EVALUATION PROCESS. BIDS MAY BE EXAMINED WITHIN 72 HOURS AFTER BID OPENING. INFORMATION PERTAINING TO COMPLETED FILES MAY BE SECURED BY VISITING THE STATE OF LOUISIANA DURING NORMAL WORKING HOURS. WRITTEN BID TABULATIONS WILL NOT BE FURNISHED.</p> <p>17. AWARDS. THE STATE OF LOUISIANA RESERVES THE RIGHT TO AWARD ITEMS SEPARATELY, GROUPED OR ON AN ALL-OR-NONE BASIS AND TO REJECT ANY OR ALL BIDS AND WAIVE ANY INFORMALITIES.</p> <p>18. PRICES . UNLESS OTHERWISE SPECIFIED BY THE STATE OF LOUISIANA IN THE SOLICITATION, BID PRICES MUST BE COMPLETE, INCLUDING TRANSPORTATION PREPAID BY BIDDER TO DESTINATION AND FIRM FOR ACCEPTANCE FOR A MINIMUM OF 30 DAYS. IF ACCEPTED, PRICES MUST BE FIRM FOR THE CONTRACTUAL PERIOD. BIDS OTHER THAN F.O.B. DESTINATION MAY BE REJECTED. PRICES SHOULD BE QUOTED IN THE UNIT (EACH, BOX, CASE, ETC.) AS SPECIFIED IN THE SOLICITATION.</p> <p>19. DELIVERIES. BIDS MAY BE REJECTED IF THE DELIVERY TIME INDICATED IS LONGER THAN THAT SPECIFIED IN THE SOLICITATION.</p> <p>20. TAXES. VENDOR IS RESPONSIBLE FOR INCLUDING ALL APPLICABLE TAXES IN THE BID PRICE. STATE AGENCIES ARE EXEMPT FROM ALL STATE AND LOCAL SALES AND USE TAXES.</p>			

STANDARD TERMS & CONDITIONS		INVITATION TO BID	
NUMBER : 2239603 OPEN DATE: 05/27/10 TIME: 10:00 AM T-NUMBER :		BIDDER:	PAGE 3
<p>21. NEW PRODUCTS.</p> <p>UNLESS SPECIFICALLY CALLED FOR IN THE SOLICITATION, ALL PRODUCTS FOR PURCHASE MUST BE NEW, NEVER PREVIOUSLY USED, AND THE CURRENT MODEL AND/OR PACKAGING. NO REMANUFACTURED, DEMONSTRATOR, USED OR IRREGULAR PRODUCT WILL BE CONSIDERED FOR PURCHASE UNLESS OTHERWISE SPECIFIED IN THE SOLICITATION. THE MANUFACTURER'S STANDARD WARRANTY WILL APPLY UNLESS OTHERWISE SPECIFIED IN THE SOLICITATION.</p> <p>22. CONTRACT RENEWALS.</p> <p>UPON AGREEMENT OF THE STATE OF LOUISIANA AGENCY AND THE CONTRACTOR, A TERM CONTRACT MAY BE EXTENDED FOR 2 ADDITIONAL 12-MONTH PERIODS AT THE SAME PRICES, TERMS AND CONDITIONS. IN SUCH CASES, THE TOTAL CONTRACT TERM CANNOT EXCEED 36 MONTHS.</p> <p>23. CONTRACT CANCELLATION.</p> <p>THE STATE OF LOUISIANA HAS THE RIGHT TO CANCEL ANY CONTRACT, IN ACCORDANCE WITH PURCHASING RULES AND REGULATIONS, FOR CAUSE, INCLUDING BUT NOT LIMITED TO, THE FOLLOWING: (1) FAILURE TO DELIVER WITHIN THE TIME SPECIFIED IN THE CONTRACT; (2) FAILURE OF THE PRODUCT OR SERVICE TO MEET SPECIFICATIONS, CONFORM TO SAMPLE QUALITY OR TO BE DELIVERED IN GOOD CONDITION; (3) MISREPRESENTATION BY THE CONTRACTOR; (4) FRAUD, COLLUSION, CONSPIRACY OR OTHER UNLAWFUL MEANS OF OBTAINING ANY CONTRACT WITH THE STATE; (5) CONFLICT OF CONTRACT PROVISIONS WITH CONSTITUTIONAL OR STATUTORY PROVISIONS OF STATE OR FEDERAL LAW; (6) ANY OTHER BREACH OF CONTRACT.</p> <p>24. DEFAULT OF CONTRACTOR.</p> <p>FAILURE TO DELIVER WITHIN THE TIME SPECIFIED IN THE BID WILL CONSTITUTE A DEFAULT AND MAY CAUSE CANCELLATION OF THE CONTRACT. WHERE THE STATE HAS DETERMINED THE CONTRACTOR TO BE IN DEFAULT, THE STATE RESERVES THE RIGHT TO PURCHASE ANY OR ALL PRODUCTS OR SERVICES COVERED BY THE CONTRACT ON THE OPEN MARKET AND TO CHARGE THE CONTRACTOR WITH COST IN EXCESS OF THE CONTRACT PRICE. UNTIL SUCH ASSESSED CHARGES HAVE BEEN PAID, NO SUBSEQUENT BID FROM THE DEFAULTING CONTRACTOR WILL BE CONSIDERED.</p> <p>25. ORDER OF PRIORITY.</p> <p>IN THE EVENT THERE IS A CONFLICT BETWEEN THE INSTRUCTIONS TO BIDDERS OR STANDARD CONDITIONS AND THE SPECIAL CONDITIONS, THE SPECIAL CONDITIONS SHALL GOVERN.</p> <p>26. APPLICABLE LAW.</p> <p>ALL CONTRACTS SHALL BE CONSTRUED IN ACCORDANCE WITH AND GOVERNED BY THE LAWS OF THE STATE OF LOUISIANA.</p> <p>27. COMPLIANCE WITH CIVIL RIGHTS LAWS.</p> <p>BY SUBMITTING AND SIGNING THIS BID, BIDDER AGREES TO ABIDE BY THE REQUIREMENTS OF THE FOLLOWING AS APPLICABLE: TITLE VI AND VII OF THE CIVIL RIGHTS ACT OF 1964, AS AMENDED BY THE EQUAL OPPORTUNITY ACT OF 1972, FEDERAL EXECUTIVE ORDER 11246, FEDERAL REHABILITATION ACT OF 1973, AS AMENDED, THE VETERAN'S READJUSTMENT ASSISTANCE ACT OF 1974, TITLE IX OF THE EDUCATION AMENDMENTS OF 1972, THE AGE ACT OF 1975, AND BIDDER AGREES TO ABIDE BY THE REQUIREMENTS OF THE AMERICANS WITH DISABILITIES ACT OF 1990. BIDDER AGREES NOT TO DISCRIMINATE IN ITS EMPLOYMENT PRACTICES, AND WILL RENDER SERVICES UNDER ANY CONTRACT ENTERED INTO AS A RESULT OF THIS SOLICITATION WITHOUT REGARD TO RACE, COLOR, RELIGION, SEXUAL ORIENTATION, NATIONAL ORIGIN, VETERAN STATUS, POLITICAL AFFILIATION, OR DISABILITIES. ANY ACT OF DISCRIMINATION COMMITTED BY BIDDER, OR FAILURE TO COMPLY WITH THESE STATUTORY OBLIGATIONS WHEN APPLICABLE, SHALL BE GROUNDS FOR TERMINATION OF ANY CONTRACT ENTERED INTO AS A RESULT OF THIS SOLICITATION.</p> <p>28. SPECIAL ACCOMMODATION.</p> <p>ANY "QUALIFIED INDIVIDUAL WITH A DISABILITY" AS DEFINED BY THE AMERICANS WITH DISABILITIES ACT WHO HAS SUBMITTED A BID AND DESIRES TO ATTEND THE BID OPENING, MUST NOTIFY THIS OFFICE IN WRITING NOT LATER THAN SEVEN DAYS PRIOR TO THE BID OPENING DATE OF THEIR NEED FOR SPECIAL ACCOMMODATIONS. IF THE REQUEST CANNOT BE REASONABLY PROVIDED, THE INDIVIDUAL WILL BE INFORMED PRIOR TO THE BID OPENING.</p> <p>29. INDEMNITY.</p> <p>CONTRACTOR AGREES, UPON RECEIPT OF WRITTEN NOTICE OF A CLAIM OR ACTION, TO DEFEND THE CLAIM OR ACTION, OR TAKE OTHER APPROPRIATE MEASURE, TO INDEMNIFY, AND HOLD HARMLESS, THE STATE, ITS OFFICERS, ITS AGENTS AND ITS EMPLOYEES FROM AND AGAINST ALL CLAIMS AND ACTIONS FOR BODILY INJURY, DEATH OR PROPERTY DAMAGES CAUSED BY THE FAULT OF THE CONTRACTOR, ITS OFFICERS, ITS AGENTS, OR ITS EMPLOYEES. CONTRACTOR IS OBLIGATED TO INDEMNIFY ONLY TO THE EXTENT OF THE FAULT OF THE CONTRACTOR, ITS OFFICERS, ITS AGENTS, OR ITS EMPLOYEES. HOWEVER, THE CONTRACTOR SHALL HAVE NO OBLIGATION AS SET FORTH ABOVE WITH RESPECT TO ANY CLAIM OR ACTION FROM BODILY INJURY, DEATH OR PROPERTY DAMAGES ARISING OUT OF THE FAULT OF THE STATE, ITS OFFICERS, ITS AGENTS OR ITS EMPLOYEES.</p> <p>30. SIGNATURE AUTHORITY.</p> <p>IN ACCORDANCE WITH L.R.S. 39:1594 (ACT 121), THE PERSON SIGNING THE BID MUST BE:</p> <ol style="list-style-type: none"> 1. A CURRENT CORPORATE OFFICER, PARTNERSHIP MEMBER OR OTHER INDIVIDUAL SPECIFICALLY AUTHORIZED TO SUBMIT A BID AS REFLECTED IN THE APPROPRIATE RECORDS ON FILE WITH THE SECRETARY OF STATE; OR 2. AN INDIVIDUAL AUTHORIZED TO BIND THE VENDOR AS REFLECTED BY A CORPORATE RESOLUTION, CERTIFICATE OR AFFIDAVIT; OR 3. OTHER DOCUMENTS INDICATING AUTHORITY WHICH ARE ACCEPTABLE TO THE PUBLIC ENTITY. 			

SPECIAL TERMS & CONDITIONS		INVITATION TO BID	
NUMBER : 2239603 OPEN DATE: 05/27/10 TIME: 10:00 AM T-NUMBER :		BIDDER:	PAGE 4

1

SPECIFICATIONS: COULTER LH 500-HEMOTOLOGY ANALYZER

THE INSTRUMENT MUST MEET OR EXCEED THE FOLLOWING REQUIREMENTS:

IDENTIFICATION, SAMPLE HANDLING, AND ASPIRATION

SAMPLE HANDLING

HAVE AUTOMATED SAMPLE HANDLING, INCLUDING MIXING AND CAP-PIERCING CAPABILITY FOR AT LEAST 25 SAMPLES.

MUST BE ABLE TO AUTOMATICALLY MIX, SAMPLE AND ANALYZE CALIBRATION AND CONTROL MATERIALS.

FAIL-SAFE SAMPLE MANAGEMENT

MUST HAVE FAIL-SAFE SAMPLE MANAGEMENT SYSTEM.

POSITIVE TUBE IDENTIFICATION (TUBE BAR-CODE READ AT TIME OF ASPIRATION)

POSITIVE CASSETTE/POSITION IDENTIFICATION (CASSETTE AND POSITION BAR-CODE READ AT TIME OF ASPIRATION)

SAMPLE VOLUME

REQUIRE 185 UL WHOLE BLOOD FOR ANALYSIS IN THE AUTOMATED (CAP-PIERCING) MODE

REQUIRE 125 UL SAMPLE IN THE MANUAL OR PRE-DILUTE MODE OF A 1:3 DILUTION (MINIMUM 50 UL WHOLE BLOOD)

REQUIRE 50 UL WHOLE BLOOD WITH 4 DROPS REAGENT A AND 2ML REAGENT B FOR RETIC MODE. AUTOMATIC PROBE WIPE AND WASH IN MANUAL MODE.

DUAL SAMPLE ASPIRATION VERIFIERS

SPECIMEN INTEGRITY CHECK DURING AND AFTER ASPIRATION.

POSITIVE BARCODE ID

MUST MEET THESE SPECIFICATIONS FOR BAR CODE READERS AND LABELS

CODE 39
CODABAR
NW7
INTERLEAVED 2 OF 5
CODE 128

HAVE 2 IDENTIFICATION FIELDS, EACH OF WHICH CAN BE ALPHA NUMERIC OR ALPHA OR NUMERIC UP TO 16 CHARACTERS

HAVE A BAR CODED CASSETTE NUMBER AND POSITION FOR ADDITIONAL AUDIT TRAIL ALLOWING OPERATOR TO EASILY IDENTIFY THE TUBE IN FURTHER REVIEW IS NECESSARY

SPECIAL TERMS & CONDITIONS		INVITATION TO BID	
NUMBER : 2239603 OPEN DATE: 05/27/10 T-NUMBER :	TIME: 10:00 AM	BIDDER:	PAGE 5

COUNTING, DIFFERENTIAL AND FLAGGING TECHNOLOGY

MUST EXTEND ANALYSIS TIME FOR CYTOPENIC SAMPLES (RBC, PLT AND WBC)

MUST DIRECTLY MEASURE MCV

MUST HAVE VISUAL DISPLAYS RBC VOLUME DISTRIBUTION CURVES OVER THE RANGE 24 - 360 FL USING 256 CHANNELS WHICH ARE 1.31 FL IN SIZE TO PROVIDE MAXIMUM RESOLUTION

MUST HAVE VISUAL DISPLAYS OF PLATELET VOLUME DISTRIBUTION CURVES FROM 2 - 20 FL USING 64 CHANNELS WHICH ARE 0.28 FL IN SIZE TO PROVIDE MAXIMUM RESOLUTION

MUST USE SWEEP FLOW TO MAXIMIZE SENSING IN THE RBC/PLT APERTURES

MUST PROVIDE A DIFFERENTIAL ANALYSIS IN A SINGLE CHANNEL USING THE THREE TECHNOLOGIES OF VOLUME, CONDUCTIVITY, AND SCATTER IN ORDER TO MAXIMIZE RESOLUTION, SPECIFICITY, AND EFFICIENCY

MUST BE ABLE TO PROVIDE A DIFFERENTIAL ON WBC EQUAL OR GREATER THAN 100 CELLS/UL

MUST DIRECTLY MEASURE EACH LEUKOCYTE POPULATION (VS. CALCULATED PARAMETERS) IN ORDER TO PROVIDE MORE ACCURATE RESULTS

MUST ANALYZE LEUKOCYTES IN THEIR NEAR-NATIVE STATE SO AS NOT TO PHYSICALLY OR BIOCHEMICALLY ALTER THEIR MORPHOLOGY. (NO CYTOCHEMICAL STAIN, THEREFORE A MUCH CLEANER SYSTEM AND MORE RELIABLE RESULTS)

MUST INCLUDE CUSTOMER DIFINABLE DEFINITIVE FLAGS AND INSTRUMENT GENERATED SUSPECT FLAGS
MODE OF OPERATION

MUST BE CAPABLE OF AUTOMATIC, CAP-PIERCING WHOLE BLOOD MODE RUNNING BY PROFILE TYPE:
CBC
CBC/DIFF

PARAMETERS:

PERFORM AND DISPLAY A HEMATOLOGY PROFILE CONSISTING OF AT LEAST THE FOLLOWING 26 PARAMETERS:

WHITE BLOOD CELL COUNT PER UL
LYMPHOCYTE %
MONOCYTE %
NEUTROPHIL %
EOSINOPHIL %
BASOPHIL %
LYMPHOCYTE #
MONOCYTE #
NEUTROPHIL #
EOSINOPHIL #
BOSOPHIL #
RED BLOOD CELL COUNT PER UL

SPECIAL TERMS & CONDITIONS		INVITATION TO BID	
NUMBER : 2239603 OPEN DATE: 05/27/10 TIME: 10:00 AM T-NUMBER :		BIDDER:	PAGE 6

HEMOGLOBIN G/DL
 HEMATOCRIT %
 MEAN CORPUSCULAR VOLUME FL
 MEAN CORPUSCULAR HEMOGLOBIN PG
 MEAN CORPUSCULAR HEMOGLOBIN CONCENTRATION G/DL
 RED CELL DISTRIBUTION WIDTH %
 PLATELET COUNT PER UL
 MEAN PLATELET VOLUME FL
 PLATELETCRIT %
 PLATELET DISTRIBUTION WIDTH %
 RETICULOCYTE #
 RETICULOCYTE %
 MEAN RETICULOCYTE VOLUME
 IMMATURE RETICULOCYTE FRACTION

DATA MANAGEMENT

MUST PROVIDE AN INFORMATION MANAGEMENT SYSTEM OFFERING FLEXIBILITY FOR EASY INTERFACING, ADDITION OF PERIPHERAL DEVICES AND FUTURE SOFTWARE VIA DISK

INFORMATION MANAGEMENT

INFORMATION MANAGEMENT SYSTEM MUST BE ABLE TO PROVIDE:

3 DIFFERENT LEVELS OF USER ACCESS FOR SECURITY OF DATA AND SYSTEM CONFIGURATION

AT LEAST 20,000 NUMERIC PATIENT RESULTS, INCLUDING GRAPHIC RESULTS AND LIST MODE DATA

AN UNLIMITED NUMBER OF USER-DEFINABLE CONTROL FILES

AN UNLIMITED NUMBER OF PATIENT CONTROL FILES

SETUP OF FILES FOR EACH SHIFT WORKED

RECORD REAGENT INFORMATION LOT NUMBERS ETC.

ENTRY OF CONTROL DATA FROM DISKETTE

AUTOSTOP IN THE EVENT OF UNACCEPTABLE CONTROL DATA

A DATABASE OF LOCATION AND PHYSICIAN

USER DEFINED DECISION RULES TO OPTIMIZE AND STANDARDIZE LABORATORY PROCESSES 24 HRS A DAY

TRANSMISSION OF PATIENT RESULTS, CONTROL RESULTS, NUMERIC AND GRAPHIC TO AS HOST COMPUTER.

SHORT TERM STORAGE

MUST PROVIDE SHORT TERM STORAGE FOR:

SPECIAL TERMS & CONDITIONS		INVITATION TO BID	
NUMBER : 2239603 OPEN DATE: 05/27/10 TIME: 10:00 AM T-NUMBER :		BIDDER:	PAGE 7
<div> <div></div> <div> <p> STARTUP LOG CALIBRATION REPRODUCTIBILITY CARRYOVER MAINTENANCE SYSTEM EVENTS </p> <p>QC</p> <p>COMBINE THE FOLLOWING QC TECHNIQUES:</p> <p>STORAGE AND ANALYSIS OF COMMERCIAL CONTROL DATA</p> <p>DAILY INSTRUMENT CHECK</p> <p>PATIENT RESULTS BY OPERATOR</p> <p>STORAGE OF PATIENT RESULTS WITH XB ANALYSIS WITH AUTOSHOP FUNCTION AFTER TWO SUCCESSIVE BATCHES EXCEED LIMITS</p> <p>ARCHIVE ANY OR ALL CONTROL DATA TO FLOPPY DISK</p> <p>INTERLABORATORY QUALITY ASSURANCE PROGRAM (IQAP)</p> <p>IQAP ASSESSMENT</p> <p>AUTOMATED CALIBRATION WITH VALUES FROM DISKETTE</p> <p>HOST COMMUNICATION</p> <p>MUST HAVE BI-DIRECTINAL COMMUNICATION CAPABILITY WITH LABORATORY INFORMATION SYSTEM</p> <p>MUST BE ABLE TO ACCEPT PRE-ASSIGNED WORKLIST FROM LIS INCLUDING PATIENT DEMOGRAPHICS</p> <p>MUST OFFER A CHOICE OF PRIMARY SAMPLE IDENTIFIERS</p> <p>MUST BE ABLE TO EDIT RESULTS AND ADD APPROPRIATE COMMENTS</p> <p>MUST TRANSMIT NUMERICAL SAMPLE RESULTS, FLAGS, GRAPHICS AND QUALITY CONTROL RESULTS TO HOST</p> <p>MAINTENANCE</p> <p>HAVE ZERO ROUTINE DAILY MAINTENANCE WHICH INCLUDES UTILIZATION OF THE FOLLOWING FEATURES:</p> <p>SELF-CLEANING BLOOD SAMPLING VALVE</p> <p>PROPRIETARY PULL-APART PINCH TUBING</p> <p>CLEANING AGENT, WHICH ELIMINATES THE NEED FOR ROUTINE BLEACHING</p> <p>BURN CIRCUIT, WHICH PURGES APERTURES OF SAMPLE RESIDUES AFTER EACH</p> </div> </div>			

SPECIAL TERMS & CONDITIONS		INVITATION TO BID																										
NUMBER : 2239603 OPEN DATE: 05/27/10 TIME: 10:00 AM T-NUMBER :		BIDDER:	PAGE 8																									
<p> CYCLE WITH SPECIFIED VOLTAGE COULTER CLENZ CLEANING AGENT AUTOMATIC ANTI-CLOG DEVICE FOR VCS CELL ANALYSIS CHAMBER MUST HAVE STAT CAPABILITY WITH POSITIVE BAR CODE IDENTIFICATION CAP PIERCING NEEDLE SHOULD BE SINGLE NEEDLE ASSEMBLY, WHICH BOTH PIERCES AND VENTS WHILE MINIMIZING TRAUMA TO THE RUBBER CAP MUST PROVIDE SENSING ALERTS FOR LOW REAGENTS MANUFACTURER MUST PROVIDE COMMERCIAL CONTROL MATERIALS FOR ROUTINE QUALITY CONTROL OF CBC AND EXTENDED DIFFERENTIAL PARAMETERS REQUIRES A MAXIMUM OF 5 REAGENTS INCLUDING CLEANING AGENT IN ORDER TO MINIMIZE INVENTORY, LOT-TO-LOT QUALITY CONTROL, MAINTENANCE AND CALIBRATION PHYSICAL/INSTALLATION SPECIFICATIONS DIMENSIONS/WEIGHT </p> <table> <thead> <tr> <th>COMPONENT</th> <th>HEIGHT</th> <th>WIDTH</th> <th>DEPTH</th> <th>WEIGHT</th> </tr> </thead> <tbody> <tr> <td>LH 500 HEMATOLOGY ANALYZER</td> <td>MAXIMUM 84.84 CM (33.40IN.)</td> <td>61.5 CM (25.10 IN)</td> <td>61.7 CM (25.20 IN.)</td> <td>94.5 K (210 LB)</td> </tr> <tr> <td>COMPUTER(S)</td> <td>43.18 CM (17.0 IN.)</td> <td>19.7 CM (7.75 IN.)</td> <td>47.6 CM (18.75 IN.)</td> <td>9.7 K (21.5 LB)</td> </tr> <tr> <td>LCD DISPLAY</td> <td>41.9 CM (16.5 IN.)</td> <td>40.4 CM (15.9 IN.)</td> <td>19.6 CM (7.7 IN.)</td> <td>7.34 K (16.3 LB.)</td> </tr> <tr> <td>KEYBOARD</td> <td>3.6 CM (1.4 IN.)</td> <td>40.4 CM (15.9 IN.)</td> <td>18.0 CM (7.1 IN.)</td> <td>2.8 K (6.2 LB)</td> </tr> </tbody> </table> <p> POWER/INPUT/CONSUMPTION CATEGORY II (PER IEC 1010-1). 100/120 VAC 50/60 HZ. 220/240 VAC 50/60 HZ. 1440 WATTS INCLUDING ALL PERIPHERALS ENVIRONMENT TEMPERATURE, AMBIENT OPERATING HUMIDITY 16 TO 320 F) 0 TO 95% NON CONDENSING PERFORMANCE SPECIFICATIONS THROUGHPUT </p>				COMPONENT	HEIGHT	WIDTH	DEPTH	WEIGHT	LH 500 HEMATOLOGY ANALYZER	MAXIMUM 84.84 CM (33.40IN.)	61.5 CM (25.10 IN)	61.7 CM (25.20 IN.)	94.5 K (210 LB)	COMPUTER(S)	43.18 CM (17.0 IN.)	19.7 CM (7.75 IN.)	47.6 CM (18.75 IN.)	9.7 K (21.5 LB)	LCD DISPLAY	41.9 CM (16.5 IN.)	40.4 CM (15.9 IN.)	19.6 CM (7.7 IN.)	7.34 K (16.3 LB.)	KEYBOARD	3.6 CM (1.4 IN.)	40.4 CM (15.9 IN.)	18.0 CM (7.1 IN.)	2.8 K (6.2 LB)
COMPONENT	HEIGHT	WIDTH	DEPTH	WEIGHT																								
LH 500 HEMATOLOGY ANALYZER	MAXIMUM 84.84 CM (33.40IN.)	61.5 CM (25.10 IN)	61.7 CM (25.20 IN.)	94.5 K (210 LB)																								
COMPUTER(S)	43.18 CM (17.0 IN.)	19.7 CM (7.75 IN.)	47.6 CM (18.75 IN.)	9.7 K (21.5 LB)																								
LCD DISPLAY	41.9 CM (16.5 IN.)	40.4 CM (15.9 IN.)	19.6 CM (7.7 IN.)	7.34 K (16.3 LB.)																								
KEYBOARD	3.6 CM (1.4 IN.)	40.4 CM (15.9 IN.)	18.0 CM (7.1 IN.)	2.8 K (6.2 LB)																								

SPECIAL TERMS & CONDITIONS		INVITATION TO BID	
NUMBER : 2239603 OPEN DATE: 05/27/10 T-NUMBER :	TIME: 10:00 AM	BIDDER:	PAGE 9

PLT>300X10 3 CELLS/UL

LINEARITY

LINEARITY, WHEN TESTED USING A STABLE SAMPLE HAVING NO INTERFERING SUBSTANCES

PARAMETER	LINEARITY RANGE	DIFFERENCE (WHICHEVER IS GREATER)
WBC	0.0-20.0 X 10 3 CELLS/UL	+0.2 OR 3.5%
	20.1-200 X 10 3 CELLS/UL	4.1%
RBC	0.0-7.00 X 10 6 CELLS/UL	+0.05 OR 2.0%
HGB	0.0-25.0 G/DL	+0.2 OR 2.0%
MCV	50-150.0 FL	+2.0 OR 2.0%
PLT	-2000 X 10 3 CELLS/UL	+10.0 OR 7.0%
MPV	5.0 TO 20.0 FI	+5.0%

PRECISION

PRECISION IS BASED ON REPLICATE DETERMINATION OF THE SAME SAMPLE:

PARAMETER	RANGE	%CV
WBC	4.0-15.0X10 3 CELLS/UL	>2.5
RBC	3.00-6.00X10 6 CELLS/UL	>2.0
HGB	12.0-18.0 G/DL	>1.5
MCV	80.0-100.0 FL	>2.0
PLT	200-500X10 3 CELL/UL	>5.0
MPV	7.0-12.0 FL	>3.0
PARAMETER	RANGE	2SD
LY%	25-35%	+3.0
MO%	5-10%	+2.0
NE%	50-60%	+3.0
EO%	2-5%	+1.0
BA%	0.5-1.5%	+1.0
PARAMETER	RANGE	%CV
RET%	<1.00%	<23% OR <0.23SD
RET%	1.00 - 4.00%	<17% OR <0.23SD
RET%	4.01-15.00%	<15% OR <0.68SD

ACCURACY

AGREEMENT BETWEEN TEST AND REFERENCE VALUES AT ANY POINT WITHIN THE OPERATING RANGE

PARAMETER	UNITS	RANGE	MEAN DIFFERENCE (THE GREATER OF)
WBC	X 10 3 CELLS/UL	0 - 20	+0.2 OR 3.5%
RBC	X 10 6 CELLS/UL	0 - 7	+0.05 OR 2.0%
HGB	G/DL	0 - 25	+0.2 OR 3.0%
MCV	FL	50-150	+2.0%
PLT	X 10 3 CELLS/UL	0 - 2000	+10.0 OR 7.0%
MPV	FL	5 - 20	+5.0%
RET	%		+5

PARAMETER	MEAN DIFF USING LH750 ANALYZER
NE	+2.0

SPECIAL TERMS & CONDITIONS		INVITATION TO BID	
NUMBER : 2239603 OPEN DATE: 05/27/10 TIME: 10:00 AM T-NUMBER :		BIDDER:	PAGE 10

	LY	+_1.5	
	MO	+_1.5	
	EO	+_0.5	
	BA	+_0.5	

BACKGROUND

PARAMETER	UNITS	RESULT
WBC	X 10 3 CELLS/UL	<0.40
RBC	X 10 6 CELLS/UL	<0.04
HGB	G/DL	<0.2
PLT	X 10 3 CELLS/UL	<3.00

MODE TO MODE

PARAMETER	UNITS	LIMITS: (THE GREATER OF)
WBC	X 10 3 CELL/UL	+_0.4 OR 5.0%
RBC	X 10 6 CELL/UL	+_0.2 OR 2.0%
HGB	G/DL	+_0.3 OR 2.0%
PLT	X 10 3 CELLS/UL	+_20.0 OR 7.0%

CLINICAL SENSITIVITY

CRITERIA	NEGATIVE	POSTIVE	TOTAL
REFERENCE NORMAL	177 (TN)	26 (FP)	FP = 12.8%
REFERENCE ABNORMAL	1 (FN)	26 (TP)	FN + 3.7%
TOTAL	178	52	230
EFFICIENCY			88.2%
REVIEW RATE			11.8%

$\% \text{EFFICIENCY} = (TP + TN) / \text{TOTAL} * 100.$ $\% \text{FP} = 100 / (TN + FP) * \text{FP}.$ $\% \text{FN} = 100 / (TP + FN) * \text{FN}$

EQUIPMENT:

LH 500 HEMATOLOGY SYSTEM INCLUDES INSTALLATION OF THE LH 500 ANALYZER WITH RETICULOCYTE CAPABILITIES, ONE PROSERVICE REMOTE DIAGNOSTIC CONNECTION, A WINDOW-BASED WORKSTATION, FLAT SCREEN MONITOR, PRINTER, UPS, INTEGRATED BARCODE SCANNER, PLUS AN EXTRA HAND-HELD BARCODE SCANNER AND TRAINING FOR ONE PERSON IN BECKMAN COULTER'S EDUCATION CENTER IN MIAMI, FLORIDA. AIRFARE WILL BE PROVIDED AS PART OF THE PRICE OF THE EQUIPMENT.

NO MONTHLY PAYMENT FOR THE LEASE OF THE EQUIPMENT WILL BE MADE UNTIL JULY 1, 2010 THUR JUNE 30, 2010 AND THEN FOR A EXTRA 2 MORE YEARS IF VILLA DECIDES TO RENEW FOR THE NEXT YEAR.

2 BID DELIVERY INSTRUCTIONS FOR STATE PURCHASING:

BIDDERS ARE HEREBY ADVISED THAT THE U.S. POSTAL SERVICE DOES NOT MAKE DELIVERIES TO OUR PHYSICAL LOCATION.

BIDS MAY BE MAILED THROUGH THE U.S. POSTAL SERVICE TO OUR BOX AT:

OFFICE OF STATE PURCHASING
 P O BOX 94095
 BATON ROUGE LA 70804-9095

SPECIAL TERMS & CONDITIONS		INVITATION TO BID	
NUMBER : 2239603 OPEN DATE: 05/27/10 TIME: 10:00 AM T-NUMBER :		BIDDER:	PAGE 11
<p>BIDS MAY BE DELIVERED BY HAND OR COURIER SERVICE TO OUR PHYSICAL LOCATION AS FOLLOWS:</p> <p>OFFICE OF STATE PURCHASING CLAIBORNE BUILDING, SUITE 2-160 1201 NORTH THIRD STREET BATON ROUGE, LA 70802</p> <p>BIDDER IS SOLELY RESPONSIBLE FOR ENSURING THAT ITS COURIER SERVICE PROVIDER MAKES INSIDE DELIVERIES TO OUR PHYSICAL LOCATION. THE OFFICE OF STATE PURCHASING IS NOT RESPONSIBLE FOR ANY DELAYS CAUSED BY THE BIDDER'S CHOSEN MEANS OF BID DELIVERY.</p> <p>BIDDER SHOULD BE AWARE OF SECURITY REQUIREMENTS FOR THE CLAIBORNE BUILDING AND ALLOW TIME TO BE PHOTOGRAPHED AND PRESENTED WITH A TEMPORARY IDENTIFICATION BADGE.</p> <p>BIDDER IS SOLELY RESPONSIBLE FOR THE TIMELY DELIVERY OF ITS BID. FAILURE TO MEET THE BID OPENING DATE & TIME SHALL RESULT IN REJECTION OF THE BID.</p> <p>***** PUBLICIZING AWARDS. IN ACCORDANCE WITH L.A.C.34:I.535, UNSUCCESSFUL BIDDERS WILL BE NOTIFIED OF THE AWARD PROVIDED THAT THEY SUBMIT WITH THEIR BID A SELF-ADDRESSED STAMPED ENVELOPE REQUESTING THIS INFORMATION. *****</p> <p>**ATTENTION:**</p> <p>RECEIPT OF A SOLICITATION OR AWARD CANNOT BE RELIED UPON AS AN ASSURANCE OF RECEIVING FUTURE SOLICITATIONS. IN ORDER TO RECEIVE FUTURE SOLICITATIONS/AWARDS FROM THIS OFFICE, YOU MUST ENROLL IN THE PROPER CATEGORY ON LAPAC AT THE FOLLOWING WEB SITE: HTTP://WWWPRD.DOA.LOUISIANA.GOV/OSP/LAPAC/PUBMAIN.ASP ENROLLMENT IN LAPAC IS FREE AND PROVIDES EMAIL NOTIFICATION OF BID OPPORTUNITIES BASED UPON COMMODITIES THAT YOU SELECT.</p> <p>3 TERMS AND CONDITIONS. THIS SOLICITATION CONTAINS ALL TERMS AND CONDITIONS WITH RESPECT TO THE COMMODITIES HEREIN. ANY VENDOR CONTRACTS, FORMS, TERMS OR OTHER MATERIALS SUBMITTED WITH BID MAY CAUSE BID TO BE REJECTED.</p> <p>4 VENDOR'S FORMS. THE PURCHASE/RELEASE ORDER IS THE ONLY BINDING DOCUMENT TO BE ISSUED AGAINST THIS CONTRACT. SIGNING OF VENDOR'S FORMS IS NOT ALLOWED.</p> <p>5 COMPLIANCE WITH CIVIL RIGHTS LAWS. BY SUBMITTING AND SIGNING THIS SOLICITATION, THE BIDDER AGREES TO ABIDE BY THE REQUIREMENTS OF THE FOLLOWING AS APPLICABLE: TITLE VI AND TITLE VII OF THE CIVIL RIGHTS ACT OF 1964, AS AMENDED BY THE EQUAL OPPORTUNITY ACT OF 1972, FEDERAL EXECUTIVE ORDER 11246, THE FEDERAL REHABILITATION ACT OF 1973, AS</p>			

SPECIAL TERMS & CONDITIONS		INVITATION TO BID	
NUMBER : 2239603 OPEN DATE: 05/27/10 TIME: 10:00 AM T-NUMBER :		BIDDER:	PAGE 12

AMENDED, THE VIETNAM ERA VETERAN'S READJUSTMENT ASSISTANCE ACT OF 1974, TITLE IX OF THE EDUCATION AMENDMENTS OF 1972, THE AGE ACT OF 1975, AND BIDDER AGREES TO ABIDE BY THE REQUIREMENTS OF THE AMERICANS WITH DISABILITIES ACT OF 1990. BIDDER AGREES NOT TO DISCRIMINATE IN ITS EMPLOYMENT PRACTICES, AND WILL RENDER SERVICES UNDER ANY CONTRACT ENTERED INTO AS A RESULT OF THIS SOLICITATION, WITHOUT REGARD TO RACE, COLOR, RELIGION, SEX, NATIONAL ORIGIN, VETERAN STATUS, POLITICAL AFFILIATION, OR DISABILITIES. ANY ACT OF DISCRIMINATION COMMITTED BY BIDDER, OR FAILURE TO COMPLY WITH THESE STATUTORY OBLIGATIONS WHEN APPLICABLE, SHALL BE GROUNDS FOR TERMINATION OF ANY CONTRACT ENTERED INTO AS A RESULT OF THIS SOLICITATION.

6 PREFERENCE. IN ACCORDANCE WITH LOUISIANA REVISED STATUTES 39:1595, A PREFERENCE MAY BE ALLOWED FOR PRODUCTS MANUFACTURED, PRODUCED, GROWN, OR ASSEMBLED IN LOUISIANA OF EQUAL QUALITY.

DO YOU CLAIM THIS PREFERENCE? YES _____

SPECIFY LINE NUMBER (S) : _____

SPECIFY LOCATION WITHIN LOUISIANA WHERE THIS PRODUCT IS MANUFACTURED, PRODUCED, GROWN OR ASSEMBLED: _____

(NOTE: IF MORE SPACE IS REQUIRED, INCLUDE ON SEPARATE SHEET.)

DO YOU HAVE A LOUISIANA BUSINESS WORKFORCE? YES _____ NO _____

IF SO, DO YOU CERTIFY THAT AT LEAST FIFTY PERCENT (50%) OF YOUR LOUISIANA BUSINESS WORKFORCE IS COMPRISED OF LOUISIANA RESIDENTS?

YES _____ NO _____

FAILURE TO SPECIFY ABOVE INFORMATION MAY CAUSE ELIMINATION FROM PREFERENCES. PREFERENCES SHALL NOT APPLY TO SERVICE CONTRACTS.

7 CANCELLATION
 THE STATE OF LOUISIANA RESERVES THE RIGHT TO CANCEL THIS CONTRACT WITH THIRTY (30) DAYS WRITTEN NOTICE.

8 CERTIFICATION OF NO SUSPENSION OR DEBARMENT. BY SIGNING AND SUBMITTING ANY BID FOR \$25,000 OR MORE, THE BIDDER CERTIFIES THAT THEIR COMPANY, ANY SUBCONTRACTORS, OR PRINCIPALS ARE NOT SUSPENDED OR DEBARRED BY THE GENERAL SERVICES ADMINISTRATION (GSA) IN ACCORDANCE WITH THE REQUIREMENTS IN OMB CIRCULAR A-133.

A LIST OF PARTIES WHO HAVE BEEN SUSPENDED OR DEBARRED CAN BE VIEWED VIA THE INTERNET AT WWW.EPLS.GOV

9 CONTRACT PERFORMANCE EVALUATION
 IN AN EFFORT TO IMPROVE OUR CONTRACTS TO MEET THE NEEDS OF THE AGENCIES WE SERVE, THE CONTRACTOR'S PERFORMANCE WILL BE MONITORED.

SPECIAL TERMS & CONDITIONS		INVITATION TO BID	
NUMBER : 2239603 OPEN DATE: 05/27/10 TIME: 10:00 AM T-NUMBER :		BIDDER:	PAGE 13

 AGENCIES' FEEDBACK WILL BE REQUESTED REGARDING CUSTOMER SERVICE, DELIVERY, PRODUCT QUALITY, BILLING, OVERALL EFFECTIVENESS OF THE CONTRACT, AND ANY NEEDED CHANGES. THEIR RESPONSES WILL BE CONSIDERED IN DETERMINING OUR OPTIONS FOR RENEWALS OR REVISIONS AND REBIDDING. TO VIEW THE CONTENT OF THE CONTRACT PERFORMANCE EVALUATION FORM, GO TO [HTTP://WWW.DOA.LOUISIANA.GOV/OSP/ONLINEFORMS/SUBMIT/CONTRPERFORMANCE .PDF](http://www.doa.louisiana.gov/osp/onlineforms/submit/contrperformance.pdf), OR CALL FOR A COPY.

AGENCY REPORTS OF DEFICIENT PERFORMANCE WILL BE APPROPRIATELY ADDRESSED DURING THE CONTRACT PERIOD.

10 IF BIDDING OTHER THAN SPECIFIED, SUFFICIENT INFORMATION SHOULD BE ENCLOSED WITH THE BID IN ORDER TO DETERMINE QUALITY, SUITABILITY, AND COMPLIANCE WITH THE SPECIFICATIONS.

FAILURE TO COMPLY WITH THIS REQUEST MAY ELIMINATE YOUR BID FROM CONSIDERATION.

11 SCOPE OF CONTRACT

SUBMITTAL OF ANY TERMS AND CONDITIONS CONTRARY TO THOSE OF THE STATE OF LOUISIANA MAY CAUSE YOUR BID TO BE REJECTED. BY SIGNING BELOW, TERMS AND CONDITIONS WHICH MAY BE INCLUDED IN YOUR BID ARE NULLIFIED, AND CONTRACTOR AGREES THAT THIS CONTRACT SHALL BE CONSTRUED IN ACCORDANCE WITH AND GOVERNED BY THE LAWS OF THE STATE OF LOUISIANA.

 BIDDER SIGNATURE

12 COMPENSATION INSURANCE, PUBLIC LIABILITY, AND PROPERTY DAMAGE INSURANCE, AS OUTLINED BELOW, ARE REQUIRED IN THIS BID.

CONTRACTOR'S LIABILITY INSURANCE:

PROOF OF INSURANCE SHOULD BE SUPPLIED WITH THE BID AND WILL BE REQUIRED BEFORE WORK CAN COMMENCE.

INSURANCE COVERAGE SPECIFIED BELOW SHALL BE FURNISHED WITH THE FOLLOWING MINIMUM LIMITS:

COMPENSATION INSURANCE: THE CONTRACTOR AND SUBCONTRACTORS SHALL TAKE AND MAINTAIN DURING THE LIFE OF THE CONTRACT WORKMAN'S COMPENSATION INSURANCE FOR ALL OF THEIR EMPLOYEES EMPLOYED AT THE SITE OF THE PROJECT. IN CASE ANY CLASS OF EMPLOYEES ENGAGED IN HAZARDOUS WORK UNDER THE WORKMAN'S COMPENSATION STATUTE, THE CONTRACTOR AND SUBCONTRACTOR SHALL PROVIDE EMPLOYER'S LIABILITY INSURANCE FOR THE PROTECTION OF THEIR EMPLOYEES NOT OTHERWISE PROTECTED.

PUBLIC LIABILITY AND PROPERTY DAMAGE INSURANCE: COMPREHENSIVE PUBLIC GENERAL LIABILITY INSURANCE, INCLUDING BUT NOT LIMITED TO BODILY INJURY, PROPERTY DAMAGE, CONTRACTUAL LIABILITY, PRODUCTS LIABILITY, COMPLETED OPERATIONS AND OWNER'S PROTECTIVE LIABILITY WITH COMBINED SINGLE LIMITS OF \$500,000 PER OCCURRENCE WITH A MINIMUM AGGREGATE OF \$1,000,000.

SPECIAL TERMS & CONDITIONS		INVITATION TO BID	
NUMBER : 2239603 OPEN DATE: 05/27/10 TIME: 10:00 AM T-NUMBER :		BIDDER:	PAGE 14
<p>_____</p> <p>_____</p> <p>LICENSED AND NON-LICENSED MOTOR VEHICLES: THE CONTRACTOR SHALL TAKE OUT AND MAINTAIN DURING THE LIFE OF THE CONTRACT, AUTOMOBILE PUBLIC LIABILITY INSURANCE IN AN AMOUNT NOT LESS THAN COMBINED SINGLE LIMITS OF \$500,000 PER OCCURRENCE FOR BODILY INJURY/PROPERTY DAMAGE. IF ANY NON-LICENSED MOTOR VEHICLES ARE ENGAGED IN OPERATIONS WITHIN THE TERMS OF THE CONTRACT ON THE SITE OF THE WORK TO PERFORM THEREUNDER, SUCH INSURANCE SHALL COVER THE USE OF ALL SUCH MOTOR VEHICLES ENGAGED IN OPERATING WITHIN THE TERMS OF THE CONTRACT ON THE SITE OF THE WORK TO BE PERFORMED THEREUNDER, UNLESS SUCH COVERAGE IS INCLUDED IN THE INSURANCE SPECIFIED.</p> <p>13 ANY ORDERS RESULTING FROM THIS SOLICITATION WILL BE PAID WITH NEW FY FUNDS, IF APPROPRIATED BY THE LEGISLATURE. DELIVERY CANNOT BE MADE PRIOR TO JULY 1 AND YOUR BID PRICES MUST BE FIRM FOR ACCEPTANCE AND DELIVERY ACCORDINGLY.</p> <p>14 THE ABOVE QUANTITIES ARE ESTIMATED TO BE THE AMOUNT NEEDED. IN THE EVENT A GREATER OR LESSER QUANTITY IS NEEDED, THE RIGHT IS RESERVED BY THE STATE OF LOUISIANA TO INCREASE OR DECREASE THE AMOUNT, AT THE UNIT PRICE STATED IN THE BID.</p> <p>15 AT THE OPTION OF THE STATE OF LOUISIANA AND ACCEPTANCE BY THE CONTRACTOR, THIS CONTRACT MAY BE EXTENDED FOR TWO ADDITIONAL TWELVE (12) MONTH PERIODS AT THE SAME PRICE, TERMS AND CONDITIONS. TOTAL CONTRACT TIME MAY NOT EXCEED THIRTY-SIX (36) MONTHS.</p>			

PRICE SHEET		INVITATION TO BID			
NUMBER : 2239603 OPEN DATE : 05/27/10 TIME: 10:00 AM T-NUMBER :		BIDDER:			PAGE 15
LINE NO.	COMMODITY/SERVICE DESCRIPTION	QUANTITY FROM/TO	UNIT	UNIT PRICE	EXTENDED TOTAL
	UNLESS SPECIFIED ELSEWHERE SHIP TO: VILLA FELICIANA HOSPITAL WAREHOUSE WAREHOUSE 4988 HIGHWAY 10 JACKSON , LA 70748				
00001	COMMODITY CODE: 979-17-000000 RENTAL OF ONE COULTER LH500 HEMATOLOGY ANALYZER. SPECIFY BRAND (& NUMBER IF APPLICABLE) _____	12	MO	_____	_____
00002	COMMODITY CODE: 979-17-000000 MAINTENANCE FOR THE ABOVE ITEM ON LINE (1) TO INCLUDE ALL PARTS, LABOR, TRAVEL, 2 PREVENTIVE MAINTENCE PER YEAR. SPECIFY BRAND (& NUMBER IF APPLICABLE) _____	12	MO	_____	_____
00003	COMMODITY CODE: 193-12-000000 FP, LH SERIES DILUENT, 20L BECKMAN COULTER INC. PART # 8547194 LINES 3 THUR 11 ARE TO BE CALLED IN ON AS NEED BASES BY VILLA'S LAB DEPARTMENT. SPECIFY BRAND (& NUMBER IF APPLICABLE) _____	20	EACH	_____	_____

PRICE SHEET		INVITATION TO BID			
NUMBER : 2239603 OPEN DATE : 05/27/10 TIME: 10:00 AM T-NUMBER :		BIDDER:			PAGE 16
LINE NO.	COMMODITY/SERVICE DESCRIPTION	QUANTITY FROM/TO	UNIT	UNIT PRICE	EXTENDED TOTAL
00004	COMMODITY CODE: 193-12-000000 LYSE S III DIFF BECKMAN COULTER INC. PART# 8547189 SPECIFY BRAND (& NUMBER IF APPLICABLE) _____	1	CASE	_____	_____
00005	COMMODITY CODE: 193-12-000000 FP, LYSE A III DIFF, 1L BECKMAN COULTER INC. #8546983 SPECIFY BRAND (& NUMBER IF APPLICABLE) _____	3	EACH	_____	_____
00006	COMMODITY CODE: 193-12-000000 FP, LH700 SERIES PAK, 1.9L/0.45L BECKMAN COULTER INC. #8547195 SPECIFY BRAND (& NUMBER IF APPLICABLE) _____	6	EACH	_____	_____
00007	COMMODITY CODE: 193-12-000000 LH SERIES CLEANER, 10L BECKMAN COULTER INC. #721543	7	EACH	_____	_____

PRICE SHEET		INVITATION TO BID			
NUMBER : 2239603 OPEN DATE : 05/27/10 TIME: 10:00 AM T-NUMBER :		BIDDER:			PAGE 17
LINE NO.	COMMODITY/SERVICE DESCRIPTION	QUANTITY FROM/TO	UNIT	UNIT PRICE	EXTENDED TOTAL
_____	SPECIFY BRAND (& NUMBER IF APPLICABLE) _____				
00008	COMMODITY CODE: 193-12-000000 LATRON (LATEX) CONTROL KIT BECKMAN COULTER INC. #7546914 SPECIFY BRAND (& NUMBER IF APPLICABLE) _____	9	KIT	_____	_____
00009	COMMODITY CODE: 193-12-000000 LATRON PRIMER SOLUTION KIT BECKMAN COULTER INC. #7546915 SPECIFY BRAND (& NUMBER IF APPLICABLE) _____	9	KIT	_____	_____
00010	COMMODITY CODE: 193-12-000000 5C CELL CONTROL TRIPAK(9X3.3ML BECKMAN COULTER INC. #7547001 SPECIFY BRAND (& NUMBER IF APPLICABLE) _____	12	EACH	_____	_____
00011	COMMODITY CODE: 193-12-000000 	2	EACH	_____	_____

PRICE SHEET		INVITATION TO BID			
NUMBER : 2239603 OPEN DATE : 05/27/10 T-NUMBER :		BIDDER:			PAGE 18
LINE NO.	COMMODITY/SERVICE DESCRIPTION	QUANTITY FROM/TO	UNIT	UNIT PRICE	EXTENDED TOTAL
	FP, S-CAL, 2X4.2ML BECKMAN COULTER INC. #624519 SPECIFY BRAND (& NUMBER IF APPLICABLE)				